

[Cat. No.]      **S-6042-S200**

## Introduction

Single Gene qPCR Primer Set is designed by primer Blast (NCBI) and BIONEER's bioinformatics tool for intercalating dye-based method. All primers are designed and validated in accordance with the MIQE (Minimum Information for Publication of Quantitative Real-Time PCR Experiments) guidelines\*. The target specificity and PCR efficiency are verified and provided as a primer set of 200 reactions. This product is optimized for utilization of *AccuPower*<sup>®</sup> 2X GreenStar™ qPCR Master Mix (Cat. No. K-6251) and *Exicycler*<sup>™</sup> 96 ver. 4 (Cat. No. A-2060-1) that gives the best result. It can be used directly in the experiment without further verification, and you can derive ready-to-publish data.

\* Bustin, S.A., et al. 2009. The MIQE Guidelines: Minimum Information for Publication of Quantitative Real-Time PCR Experiments, *Clinical Chemistry* 55:4, 611-622.

## Features & Benefits

- Target-specific primer design using primer blast and our bioinformatics tool
- Exclusion of self-primer-dimer formation sequence
- Identification of single peak in the dissociation curve
- Short amplicon size of 80-160 bp
- Wide amplification range of copies of about 10<sup>2</sup>-10<sup>7</sup>
- qPCR amplification efficiency of 90-110% in compliance with the MIQE Guidelines

## Components

Components	Amount
Lyophilized Forward primer	3 nmol
Lyophilized Reverse primer	3 nmol

## Storage

- This product is lyophilized and shipped at ambient temperature. Lyophilized primers are thermo-stable, but customer can request dry-ice embedded shipment at an extra cost.
- Store at ambient temperature (15-20°C) without direct sunlight for long term storage. Once dispensed, primers should be stored at 20°C and repeated freeze and thaw cycles (more than once) are not recommended.

## Online Resources



Korean



English

Visit our **product page** for additional information and protocols

## Ordering Information

Description		Cat. No.
Single Gene qPCR Primer Set	Forward primer	200 rxn/tube
	Reverse primer	200 rxn/tube
		S-6042-S200

\* **Note:** Each of forward and reverse primer is supplied for 200 rxn in two separate single tubes.

## Notice

BIONEER corporation reserves the right to make corrections, modifications, improvements and other changes to its products, services, specifications or product descriptions at any time without notice.

## Explanation of Symbols



Caution



Consult Instructions For Use






Do not Re-use



Use-by Date

## Experimental Procedures

Steps		Procedure Details																															
1	 <b>Preparation of primers</b>	<p>1. Dissolve forward and reverse primers in 500 µl of nuclease-free water to make a concentration of 6 pmol/µl.</p>																															
2	 <b>Preparation of reaction mixture</b>	<p>2. Add template DNA, primers, nuclease-free water, and <i>AccuPower</i><sup>®</sup> 2X <i>GreenStar</i><sup>™</sup> qPCR Master Mix (K-6251, not provided) into real-time PCR plate (not provided) to make a total volume of 20 µl or 50 µl.</p> <ul style="list-style-type: none"> <li>Preparation of reaction mixture</li> </ul> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="text-align: center;">Components</th> <th style="text-align: center;">20 µl reaction</th> <th style="text-align: center;">50 µl reaction</th> </tr> </thead> <tbody> <tr> <td><i>AccuPower</i><sup>®</sup> 2X <i>GreenStar</i><sup>™</sup> qPCR Master Mix</td> <td style="text-align: center;">10 µl</td> <td style="text-align: center;">25 µl</td> </tr> <tr> <td>Template DNA</td> <td style="text-align: center;">5 pg-100 ng</td> <td style="text-align: center;">5 pg-100 ng</td> </tr> <tr> <td>Forward primer (6 pmol/µl)</td> <td style="text-align: center;">1 µl</td> <td style="text-align: center;">2.5 µl</td> </tr> <tr> <td>Reverse primer (6 pmol/µl)</td> <td style="text-align: center;">1 µl</td> <td style="text-align: center;">2.5 µl</td> </tr> <tr> <td>Nuclease-free water</td> <td style="text-align: center;">Variable</td> <td style="text-align: center;">Variable</td> </tr> <tr> <td>Total volume</td> <td style="text-align: center;">20 µl</td> <td style="text-align: center;">50 µl</td> </tr> </tbody> </table> <p>3. Seal real-time PCR plate with adhesive optical sealing film (3111-4110, not provided) and briefly spin down.</p>	Components	20 µl reaction	50 µl reaction	<i>AccuPower</i> <sup>®</sup> 2X <i>GreenStar</i> <sup>™</sup> qPCR Master Mix	10 µl	25 µl	Template DNA	5 pg-100 ng	5 pg-100 ng	Forward primer (6 pmol/µl)	1 µl	2.5 µl	Reverse primer (6 pmol/µl)	1 µl	2.5 µl	Nuclease-free water	Variable	Variable	Total volume	20 µl	50 µl										
Components	20 µl reaction	50 µl reaction																															
<i>AccuPower</i> <sup>®</sup> 2X <i>GreenStar</i> <sup>™</sup> qPCR Master Mix	10 µl	25 µl																															
Template DNA	5 pg-100 ng	5 pg-100 ng																															
Forward primer (6 pmol/µl)	1 µl	2.5 µl																															
Reverse primer (6 pmol/µl)	1 µl	2.5 µl																															
Nuclease-free water	Variable	Variable																															
Total volume	20 µl	50 µl																															
3	 <b>Real-time PCR</b>	<p>4. Perform the reaction under the following conditions.</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="text-align: center;">Step</th> <th style="text-align: center;">Temperature</th> <th style="text-align: center;">Time</th> <th style="text-align: center;">Cycles</th> </tr> </thead> <tbody> <tr> <td>Pre-denaturation</td> <td style="text-align: center;">95°C</td> <td style="text-align: center;">10 min</td> <td style="text-align: center;">1 cycle</td> </tr> <tr> <td>Denaturation</td> <td style="text-align: center;">95°C</td> <td style="text-align: center;">5 sec</td> <td></td> </tr> <tr> <td>Annealing</td> <td style="text-align: center;">58°C</td> <td style="text-align: center;">25 sec</td> <td rowspan="2" style="text-align: center;">40 cycles</td> </tr> <tr> <td>Extension</td> <td style="text-align: center;">72°C</td> <td style="text-align: center;">30 sec</td> </tr> <tr> <td>Detection</td> <td></td> <td style="text-align: center;">Scan</td> <td></td> </tr> <tr> <td>Final extension</td> <td style="text-align: center;">65°C</td> <td style="text-align: center;">5 min</td> <td style="text-align: center;">1 cycle</td> </tr> <tr> <td>Melting</td> <td style="text-align: center;">65-95°C</td> <td style="text-align: center;">1 sec</td> <td style="text-align: center;">-</td> </tr> </tbody> </table> <p>5. After the reaction, perform data analysis.</p>	Step	Temperature	Time	Cycles	Pre-denaturation	95°C	10 min	1 cycle	Denaturation	95°C	5 sec		Annealing	58°C	25 sec	40 cycles	Extension	72°C	30 sec	Detection		Scan		Final extension	65°C	5 min	1 cycle	Melting	65-95°C	1 sec	-
Step	Temperature	Time	Cycles																														
Pre-denaturation	95°C	10 min	1 cycle																														
Denaturation	95°C	5 sec																															
Annealing	58°C	25 sec	40 cycles																														
Extension	72°C	30 sec																															
Detection		Scan																															
Final extension	65°C	5 min	1 cycle																														
Melting	65-95°C	1 sec	-																														